

## Specification Sheet

Specification Number: 8017H-05
Revision: 09
Date Issued: 18-Nov-22

**1. Part Number:** 8017H-05

### 2. Photograph:



#### 3. Product Information:

Description: SteriWare® Bowl 500ml

Use: Single use bowl

Material of Construction: High Density Polyethylene (HDPE) – virgin

Conforms to FDA 21 CFR 177.1520Conforms to EU Regulations 10/2011

• Conforms to EC Regulations 1935/2004

BSE/TSE free

Method of Construction: Single piece, injection moulded

Bowl Diameter: 129mm
Bowl Width (inc, spout & handle): 180mm
Bowl Height: 70mm
Nominal Weight: 61g

Moulding Environment: Class 100,000 Cleanroom
Primary Packing Environment: Class 10,000 Cleanroom

Individually Bagged? Yes (heat sealed PE/PET bag)

Method of Sterilisation: N/A
Quantity per Box: 100

Recommended Storage Conditions: Dry and ambient temperature

Shelf Life: 5 years from the date of manufacture
ATEX: This device has not been ATEX tested



### Sampling Systems Ltd.

4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



# **Specification Sheet**

Specification Number: 8017H-05
Revision: 09
Date Issued: 18-Nov-22

**4. Change Control:** Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to

Sterivvare range please email your contact details

change.control@sampling.com

**5. Country of Origin:** Made and packed in the UK.

**6. Repeated Use:** This is a single-use, disposable sampler. Repeated use will lead to

failure.

**7. Approval:** Signed:

Name: Kuldeep Flora

**Quality Manager** 

Date: 18-Nov-22

Position: