

## **Specification Sheet**

Specification Number: 8017H-30
Revision: 06
Date Issued: 18-Nov-22

**1. Part Number:** 8017H-30

### 2. Photograph:



#### 3. Product Information:

Description: SteriWare® Bowl 3000ml

Use: Single use bowl

Material of Construction: High Density Polyethylene (HDPE) – virgin

• Conforms to FDA 21 CFR 177.1520

Conforms to EU Regulations 10/2011

Conforms to EC Regulations 1935/2004

• BSE/TSE free

Method of Construction: Single piece, injection moulded

Max. Diameter (including handles): 256mm

OD at top rim (excluding handles): 227mm

ID at top rim: 216mm

Depth of Bowl: 109mm

Nominal Weight: 170g

Moulding Environment: Class 100,000 Cleanroom
Primary Packing Environment: Class 10,000 Cleanroom

Individually Bagged? Yes (heat sealed PE/PET bag)

Method of Sterilisation: N/A

Quantity per Box: 15

Recommended Storage Conditions: Dry and ambient temperature

Shelf Life: 5 years from the date of manufacture
ATEX: This product has not been ATEX tested



### Sampling Systems Ltd.

4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



# **Specification Sheet**

Specification Number: 8017H-30
Revision: 06
Date Issued: 18-Nov-22

**4. Change Control:** Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the

SteriWare range please email your contact details to

change.control@sampling.com

**5. Country of Origin:** Made and packed in the UK.

**6. Repeated Use:** This is a single-use, disposable sampler. Repeated use will lead to

failure.

**7. Approval:** Signed:

Name: Kuldeep Flora

**Quality Manager** 

Date: 18-Nov-22

Position: