

	<h1 style="margin: 0;">Sampler Specification</h1>	Specification Number	8050P-1000
		Revision	01

Part No. 8050P-1000

Description Disposable LiquiThief – 1000mm

Material of Construction:

Tip	PP (virgin, FDA approved – CFR 177.1520)
Outer Tube	PP (virgin, FDA approved – CFR 177.1520)
Inner Rod	PP (virgin, FDA approved – CFR 177.1520)
Clutch	PP (virgin, FDA approved – CFR 177.1520)
Piston Housing	PP (virgin, FDA approved – CFR 177.1520)
Piston Seal	Silicone Rubber (virgin, FDA approved – CFR 177.2600)
Handle	PP (virgin, FDA approved – CFR 177.1520)



Method of Construction: Ultrasonic welding

Nominal Length  
 Fully closed: 991mm  
 Fully extended: 1818mm

Diameter of Outer Tube: 21mm

Diameter of hole in tip: 1.5mm

Nominal Weight of Sampler: 151g


Assembly & Packing Environment: Class 100,000 Medical Cleanroom

Individually Bagged? Yes (heat sealed PE bag)

Method of Sterilisation: None

Number of Samplers per Box: 20

Recommended Storage Conditions: Dry and ambient temperature

Specification Approved By:	Signature: 	Name: Jonathan Fenner Position: Director Date: 24-Jan-12
----------------------------	--	--